

South Australia

## **Radiation Protection and Control (Ionising Radiation) Variation Regulations 2016**

under the *Radiation Protection and Control Act 1982*

---

### **Contents**

#### **Part 1—Preliminary**

- 1 Short title
- 2 Commencement
- 3 Variation provisions

#### **Part 2—Variation of *Radiation Protection and Control (Ionising Radiation) Regulations 2015***

- 4 Variation of regulation 70—Lights and signs fitted to X-ray analysis apparatus
  - 5 Variation of regulation 92—Fixed apparatus used for medical or veterinary diagnostic radiography or by chiropractor
  - 6 Substitution of regulation 107  
107 Persons other than patient not to remain in room during fluoroscopic procedure or test procedure
- 

### **Part 1—Preliminary**

#### **1—Short title**

These regulations may be cited as the *Radiation Protection and Control (Ionising Radiation) Variation Regulations 2016*.

#### **2—Commencement**

These regulations will come into operation on the day on which they are made.

#### **3—Variation provisions**

In these regulations, a provision under a heading referring to the variation of specified regulations varies the regulations so specified.

### **Part 2—Variation of *Radiation Protection and Control (Ionising Radiation) Regulations 2015***

#### **4—Variation of regulation 70—Lights and signs fitted to X-ray analysis apparatus**

Regulation 70(3)—delete subregulation (3) and substitute:

- (3) The lights referred to in subregulations (1) and (2) must—
    - (a) be—
-

- (i) red or amber in colour; and
- (ii) fail safe; or
- (b) consist of 2 lights, each of which is on a separate circuit from the other.

**5—Variation of regulation 92—Fixed apparatus used for medical or veterinary diagnostic radiography or by chiropractor**

Regulation 92(5)—delete "Schedule 8" and substitute:

Schedule 6

**6—Substitution of regulation 107**

Regulation 107—delete the regulation and substitute:

**107—Persons other than patient not to remain in room during fluoroscopic procedure or test procedure**

A person other than the patient must not, during any fluoroscopic procedure or any test procedure, remain in the room in which the procedure is being carried out unless—

- (a) he or she has been granted permission by the person operating the apparatus; and
- (b) he or she—
  - (i) is wearing a protective apron with a shielding value of not less than 0.25 millimetres lead equivalent; or
  - (ii) is shielded by a protective screen of a kind referred to in regulation 103(2)(b).

**Note—**

As required by section 10AA(2) of the *Subordinate Legislation Act 1978*, the Minister has certified that, in the Minister's opinion, it is necessary or appropriate that these regulations come into operation as set out in these regulations.

**Made by the Governor**

with the advice and consent of the Executive Council  
on 21 January 2016

No 8 of 2016

15MSECCS050